

ADMINISTRATION OF DAMAN AND DIU,
In Charge Medical Officer,
PRIMARY HEALTH CENTER,
KACHIGAM

No.PHC-KCH/LAB-AUTO-ANA/2013-14/

Dated : /01/2014.

E-TENDER (ON LINE) NOTICE

The In Charge Medical Officer, Primary Health Center, Kachigam, Daman on behalf of President of India, invites tender for following item through on-line on <http://daman.nprocure.com> from the manufacturer / Authorized Dealers / Suppliers having valid Licence. The tender notice also available on www.nic.daman.in

Sr. No.	Description of Items	Estimated cost	EMD (in the form of FDR)	Tender Fees (Non-refundable)
1	Purchase of "Semi Auto Analyzer"	Rs.12,00,000/-	Rs. 30,000/-	Rs.2,000/-
Last date of downloading of on line tender documents : Upto 31/01/2014 by 12.00 hours.				
Last date of submission of online tender document : Upto 31/01/2014 by 15.00 hours				
On line opening of Technical Bid : on 31-01-2014 by 15.30 hours On line opening of Price Bid : If possible on 31/01/2014				
Bidders have to submit price bid in Electronic Format only on www.nprocure.com till the last date and time for submission. Price Bid in physical format shall not be accepted in any case.				
Submission of tender fees in the form of DD, EMD in the form of FDR and other supporting documents i.e. copy of valid licence from competent authority, copy of VAT/ST/ Registration and copy of PAN/TAN of Income Tax etc., and terms and conditioned duly signed in hard copy to the undersigned by RPAD/Speed Post / by hand on or before 31/01/2014 upto 15.00 hours , however, Tender Inviting Authority shall not be responsible for any postal delay.				
The tender inviting authority reserves the right to accept or reject any or all the tenders to be received without assigning any reasons thereof.				
Bidders shall have to post their queries on E-Mail address : ddmssu.idsp@nic.in on or before dated 31/01/2014 upto 10.30 hours.				
In case bidder needs any clarification or if training required for participating in online tender, they can contact the following office. “(n) Code Solution – A division, GNFC Ltd.”, 403, GNFC Info Tower, Bodakdev, Ahmedabad-380054, Gujarat (India). E-Mail-nprocure@ncode.in Fax: + 917926857321 Website : www.nprocure.com				

(Dr. D.K.Makwana)
I/c, Medical Officer,
Primary Health Center, .

ADMINISTRATION OF DAMAN AND DIU,
PRIMARY HEALTH CENTER, KACHIGAM.

No.PHC-KCH/LAB-AUTO-ANA/2013-14/
Dated : /01/2014.

**TERMS AND CONDITIONS FOR SUPPLY AND INSTALLATION OF
“Semi Auto Analyzer”
FOR PRIMARY HEALTH CENTER, KACHIGAM, NANI DAMAN**

1. The rate(s) quoted should be strictly for free delivery F.O.R. at PHC, Kachigam and will be valid and operative for supply orders issued on or before 31/3/2014 from the date of invitation of tenders and inclusive of all Taxes, installation & commissioning charges.
2. All Taxes/ Duties/ Royalties charges payable on sales/transport etc. within and / or outside the UT / State shall be payable by the supplier.
3. No extra charge for packing, forwarding and insurance etc. will be paid on the rates quoted.
4. The rates should be quoted only for the items specified in the list of requirement and should be for the items of given special mark / manufacture.
5. Rates quoted for items other than required specification/mark/manufacture may not be considered. However, indigenous manufactures may quote their own makes provided the specifications/mark/manufacture.
6. Where specification / mark/ manufacture are not specifying by this office, the rates should be quoted only for the first class and standard quality only.
7. The tenderer should specify the name of the manufacturer for the item quoted by him along with catalogue of the item.
8. The decision of the E-Tender Inviting Officer for acceptance/rejection of **SUPPLY AND INSTALLATION OF “Semi Auto Analyzer”** supplied including the decision for equivalent specifications, standard and quality etc. of **SUPPLY AND INSTALLATION OF “Semi Auto Analyzer”** shall be final.
9. The tenderer should send in advance or enclose along with technical bid an amount of Rs. 30,000/- as Earnest Money Deposit in form of Demand Draft / F.D.R. of any Schedule Bank payable at PHC, Kachigam in favour of the officer inviting tender i.e. In Charge Medical Officer, PHC, Kachigam, Daman. The EMD submitted other than form mentioned above will not be accepted. Tender received without Earnest Money Deposit will be summarily rejected.
10. (a) The successful tenderer will have to pay within 10 days from the date of demand, an amount equal to 10% of the total value of articles
(b) Non-receipt of Security Deposit within stipulated time limit will result in automatic cancellation of the order for supply without any intimation.
(c) However in case if any articles are received for which the Security Deposit may not have been deposited, the full Security Deposit as may be due from the supplier will be recovered from the bills for such articles.
11. The amount of Earnest Money paid by the successful Tenderer(s) will be adjusted against the amount of Security Deposit to be paid by the successful tenderer(s) as per condition No. 5 above.

12. All bills for amount above Rs.5000/- should be pre-receipted on a Revenue Stamp of proper value. Bills for amount exceeding Rs.5000/- not pre-receipted on Revenue Stamp of proper value will not be accepted for payment.

13. Each bill in which Sales Tax/Vat is charged must contain the following certificate on the body of the bill.

"CERTIFIED that the goods on which Sales Tax/VAT has been charged have not been exempted under the VAT/ Central Sale Tax Act or the Rules made there under and the amount charged on account of VAT/Sales Tax on these goods is not more than what is payable under the provisions of relevant Act or Rules made there under".

14. The right to accept or reject without assigning any reasons or all tenders in part or whole is reserved with the E-Tender Inviting Officer and his decision(s) on all matters relating to acceptance or rejection of the tenders as a whole or in part will be final and binding to all.

15. If the tenderer whose tender is accepted, fails to execute the supply order within stipulated time the Earnest Money Deposit of such tenders will stand forfeited to the Government.

16. In case, the supplier does not execute the supply order placed with him, the EMD of the supplier will be forfeited to the Government and the contract for the supply shall be terminated with no further liabilities on either party to the contract.

17. No separate agreement will be required to be signed by the successful tender(s) for the purpose of this contract for supply. Rates tendered/offered in response to the concerned Tender Notice shall be considered as acceptance of all above terms and conditions for supply for all legal purpose.

18. The E- Tender should be neatly typed only on letterhead carry the name of supplier and the signature of the tenderer. No overwriting correction or erasures will be considered.

19. The rates quoted should be inclusive of all taxes, duties, surcharges, cess, freight, loading, unloading, insurance, road permits, packing, (VAT and other taxes if applicable) as applicable. No extra taxes will be given.

20. All bills should be in **TRIPLICATE** and should invariably mention the number and date of supply order.

21. The Tenders and financial bid should be submitted online on www.nprocure.com in two bid system.

22. Orders once placed should be delivered within the given time period and item should be door delivered.

23. No extra charge for packing, forwarding and insurance etc. will be paid on the rates quoted.

24. The rates should be quoted only for the items specified in the list of requirement and should be for the items of given special mark/manufacture only.

25. Rates quoted for items other than the required specification / mark / manufacture will not be considered.

26. The amount of Earnest Money paid by the tenderer(s) whose tenders are not accepted will be refunded to them.

27. Only on satisfactory completion of the supply order for and on payment of all bills of the supplier, as to be admitted for payment, the amount of Security Deposit/Earnest Money will be refunded after expiry of guarantee/warranty period, if any, or any such date/period as may be mutually agreed upon.

28. The E-tender inviting officer will consider extension of time for remitting the Security Deposit as demanded. However, in case of denial to consider such extension the supplier is bound to abide by the limit given and liable to make good any loss to the Government on account of his failure to abide by the time limit.

29. (a) Railway Receipt or other transport document should be drawn in favor of the Officer Inviting Tenders.

(b) Railway Receipt or other transport document should not be send by V.P.P. or through any Bank as this being a Government Office it is not possible to clear cash demands of Post Office/Bank for delivery of R.R. or other transport documents unless we have agreed to it as a special arrangements.

(c) Railway Receipt or transport receipt should be sent to this Office by Registered Post immediately on dispatch of goods from dispatching end.

30. The supplies of **“Semi Auto Analyzer”** of inferior quality standard or of different specifications, doses/content, brand, manufacturer etc. other than that ordered specified and/or incomplete or broken articles will not be accepted. The supplier has to replace the same at his own cost and risk. Intimation of non-acceptance of any **“Semi Auto Analyzer”** will be sent to the supplier within 10 days from the date of receipt of the stores and the same will be returned to the supplier at his own cost and risk, if he so desires and intimates accordingly within 15 days from the date of dispatch of intimation of the non-acceptance. However, if no communication is received within 15 days from the date of communication the tender Inviting Officer will not be responsible for any damage, loss etc. of such rejected articles.

31. In case of failure to replace the accepted and rejected articles from supplies made, as mentioned in the conditions, the loss undergone by the Government will be recovered from the suppliers Security Deposit/Earnest Money or payment due of any bill(s) to the extent required.

32. In case of failure to **SUPPLY AND INSTALLTAION OF “Semi Auto Analyzer”** ordered for, as per conditions and within the stipulated time, the same will be obtained from the tenderer who offered next higher rates or from any other sources, as may be decided by the tender inviting Officer and the loss to the Government on account of such purchases(s) shall be recovered from the former suppliers Security Deposit/Earnest Money or bills payable. The suppliers shall have no any right to dispute with such procedure.

33. Extension of time limit for supplies shall be consider by the Tender Inviting Officer. The extension so granted may be with levy of compensation for delay in execution of supply order up to 5% of the cost of supplies ordered for at the discretion of the authority competent to grant extension of time limit provided such request is made well in time, depending upon the circumstances and such decision in the matter will be final.

34. Demurrage charges paid by the E-Tender Inviting Officer on account of delayed receipt of dispatch documents intimation will be recovered from the bills payable to the supplier.

35. If at any time after the order for **SUPPLY AND INSTALLTAION OF “Semi Auto Analyzer”** the E-Tender Inviting Officer shall for any reason whatsoever not require the whole or part of the quantity thereof as specified in the order the tender Inviting Officer shall give notice in writing of the fact to

the supplier(s) who shall have no claim to any payment of compensation what so ever on account of any profit or advantage which the supplier(s) might have derived from the supply of articles in full, but which did not derive in consequence of the full quantity of articles not having been purchased, nor shall have any claim for compensation by reasons of any alterations having been made in the original instructions which shall invoice any curtailment of the supply originally contemplated.

36. Rates should be quoted as per the forms prescribed by the department and as per the requirement asked for.

37. The Earnest Money(s)/Security Deposit(s) paid by the tender(s) earlier against any tender(s) or supply order(s) is/are not adjustable with Earnest Money or Security Deposit required by these conditions.

38. The Tenders/offers received do not confirm with the terms and conditions of this office will be summarily rejected. If any firm desires to consider exemption from payment of Earnest Money Deposit, certified copies of its Registration with D.G.S. & D. should be attached to their tenders.

39. Supplier may ensure the goods at his own cost to safeguard the delivery of such goods dispatched by him to the consignee, the department will not be responsible for the damage or pilferage of goods during transit.

40. The tenderer should attached Scanned copies of certificate of **experience in the field, valid license, proof of fulfilling the norms of ISI/ISO/WHO/GMP, CE Certified specification if any, copy of dealership letter, licence for import, PAN No, Sales Tax No.** with his/their tender. It may please be noted that the tender received without document referred above shall not be considered.

41. Rates quoted are for **SUPPLY AND INSTALLTAION OF “Semi Auto Analyzer”** to be supplied at PHC, Kachigam.

42. Tenderer should enclose along with tender an amount of Rs. 30,000/- as Earnest Money Deposit in form of Fixed Deposit Receipt of any Nationalized/ Scheduled Bank payable at Daman in favour of In Charge Medical Officer, Primary Health Cneter, Kachigam, Daman. The EMD Submitted other than Form mentioned above will not be accepted. **Tender received without EMD will be summarily rejected.**

43. The tender fee must be enclosed in demand draft in favour of undersigned with the tender documents.

44. AMC : The rates of Annual Maintenance Contract (AMC) should be for the period of three years from the expiry of Guarantee / Warranty period of one year and should written separately in the Financial bid year-wise. The decision to accept the tender with or without AMC is reserve with the Purchase Committee.

45. The tenderers should give the guaranty / warranty for the period of not less than one year from the date of installation against any manufacturing defect.

46. The tenderer may be called for a sample / demonstration of the items quoted for which they will be informed one week in advance for arranging the necessary sample / demonstration in the hospital on a suitable date and time failing which the tender will be rejected.

47. The tenderers shall be bound to give assurance for undertaking the Annual Maintenance Contract after expiry of guarantee / warranty period.

48. The successful tenderers shall be bound to provide training if any required without any extra charges during commissioning.
49. The successful tenderers should install and commissioning the hospital equipment at the site suggested by the office.
50. The tendered quantity is tentative and the actual purchase can be 15% less or more than the quantity put to tender for all items and the tenderer is bound to supply such requirement without any demur.
51. The tendering firm must be registered with the Sales Tax/ VAT department and a copy of their registration under the Sales Tax/ VAT bearing the TIN Number be provided.
52. The terms and conditions duly accepted and signed by the tenderer should be attached with the Technical Bid.
53. The tenderer should buy back old Serum Analyzer" Accumax (AX-100)
54. The Financial Bid has to be filled in prescribed format.

Signature & Designation of
Tender Inviting Officer

(Dr. D. K. Makwana)
In Charge, Medical Officer,
Primary Health center,,
Daman & Diu.

The above terms and conditions are accepted and are binding to me / us.

Place :

**Signature of Tenderers
Name of Tenderers with
seal of the firm**

Dated :

NOTE: Please return one copy of these terms and conditions duly signed with seal of firm along with the tender.

ADMINISTRATION OF DAMAN AND DIU,
In charge Medical Officer,
PRIMARY HEALTH CENTER, kachigam

Tender Notice No.PHC-KCH/LAB-AUTO-ANA/2013-14/
Dated:- 04-01-2014

(TENDER FORM (FINANCIAL BID))

**“Semi Auto Analyzer”
UNDER BUYBACK**

Sr. No.	Item	Unit	Rate of new machine.
01.	1a “Semi Auto Analyzer” (As per Technical specification quoted in the Technical bid)	One	
	1b Buy Back Price of <i>Biochemistry Analyzer</i> Accumax (AX-100)	One	
02	Net Price= (1a-1b)		
03	A.M.C. rate after completion of Guarantee / Warrantee Period:		
(a)	First Year	1a Rs	
(b)	Second Year	1a Rs	
(c)	Third Year	1a Rs	
	Total	1a Rs	

Seal

Signature and Seal of Tenderer

Date

Technical Specification

Sr. No.	Description	Unit
01	<p><u>SPECIFICATIONS OF SEMI-AUTO ANALYZER</u></p> <p>1. It should be completely OPEN system, microprocessor controlled, programmable, semi Auto Analyzer to perform routine biochemistry tests, enzyme immunoassays (with multi standard curve with 1 blank & six standards calibration</p> <p>2. It should offer more than 130 user definable chemistry parameters and more than 50 Keys tests through Direct Access Key.</p> <p>3. The unit should be able to offer 10 modes such as 1-Point Linear (End-Point), 2-Point Linear (Fixed Time), Rate A Linear (Kinetic), 1-Point Nonlinear (End-Point), 2-Point Non-linear (Fixed Time), Rate A Non-linear (Kinetic), 1-Point Sample Blank Linear (Differential), 1-Point Sample Blank Non-linear (Differential), Absorbance & Coagulation.</p> <p>4. There should be a facility to take readings with monochromatic as well as bichromatic filter selections.</p> <p>5. It should have a peltier controlled reading block with temperature programmable for Off, 25oC, 30o C & 37o C.</p> <p>6. The flowcell volume should be 20 µl or less.</p> <p>7. Flowcell with peristaltic pump should be part of the main unit.</p> <p>8. It should have facility of triple cuvettes to use both 6mm glass cuvettes and 10mm plastic cuvettes additionally in place of flowcell & there should be a READ button to take readings from 6mm glass cuvettes and 10mm plastic cuvettes.</p> <p>9. It should have 8 narrow band static interference filters with wavelength selectable from 340 - 700 nm without any use of a moving filter wheel.</p> <p>10. It should have a large 8 lines LCD display alphanumeric display and builtin full graphic printer for printing reaction curves and test results.</p> <p>11. It should require minimum reagent per test typically not more than 500 µl/ test.</p> <p>12. It should have the facility to display the actual temperature on screen.</p> <p>13. The software should be user friendly and guide the programmer step by step with special HELP & CALIB key.</p> <p>14. The instrument should also be capable of doing coagulation assays with programmable ISI value to print results in INR.</p> <p>15. It should have facility to store minimum 1000 test results in the memory. Further it should possible to recall the stored results with Patient ID or with Date or with Date & Patient ID.</p> <p>16. It should store Reagent Blank O.D., Sample Blank OD, Standard O.D., Factor, Non-linear curves</p> <p>17. It should have facility to enter Six Digit Alphanumeric Patient ID.</p> <p>18. It should have full-fledged QC menu to give - Daily QC & Monthly Levy Jennings Graphs for two controls per test.</p> <p>19. It should work with 18 V DC using External Universal SMPS</p>	one

	<p>adapter. of Max. 100W. 90-264 V, 47-63 Hz, which requires no CVT. It should be able operate using external battery.</p> <p>20. The manufacturer / supplier should have a full-fledged service force and the manufacturer should be able to supply kits locally against orders.</p>	
--	--	--

Annexure “A”

Sr.No.	Name of Item	Estimated cost	Remark
1	Semi Auto Analyzer	Rs.12,00,000/-	