

**Government Hospital,Daman
Administration of Daman and Diu,
O/O Medical Superintendent,
Government Hospital, Daman**

No.GHD/E-TENDER/2016-2017/495

Dated: 20.07.2016

E-TENDER (ON LINE) NOTICE

The Medical Superintendent, Government Hospital, Daman on behalf of president of India, invites tenders for following items through on –line on [http: Daman.nprocure.com](http://Daman.nprocure.com) from the manufacturer/Authorized Dealers/suppliers having valid License. The tender notice also available on www.nic daman.in

Sr.No.	Description Items	Estimated cost	EMD (in the form of FDR)	Tender fees (Non refundable)
01	Purchase and of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman	45,00,000/-	1,35,000/-	2000/-
Last date of downloading of on line tender documents: up to 12.08.2016 by 12.00 hours				
Last date of submission of online tender document: up to 12.08.2016 by 15.00 hours.				
On line opening of price Bid: If possible on 12.08.2016 at 16.00 hours				
Bidders have to submit price bid in Electronic format only on www.nprocure.com till the last date and time for submission. Price bid in physical formant shall not be accepted in any case.				
Submission of tender fees in the form of DD, EMD in the form of FDR and other supporting documents i.e. copy of valid license from competent authority, copy of VAT/ST Registration and copy of PAN/TAN of income Tax etc, and terms and condition duly signed in hard copy to the undersigned by RPAD/Speed post/by hand on or before 12.08.2016 upto 15.30 hours, however, Tender Inviting authority shall not be responsible for any postal delay.				
The tender inviting authority reserves the right to accept or reject any or all the tenders to be received without assigning any reasons thereof.				
Bidders shall have to post their queries on E-mail address: ghhdmn@gmail.com on or before dated 12.08.2016 upto 11.00 hours.				
In case bidder needs any clarification or if training required for participating in online tender, they can contact the following office. (n) Code Solution-A division,GNFC Ltd” 403,GNFC Info Tower,Bodakdev, Ahmedabad- 380054, Gujarat (India) E-Mail nprocure @ncode.in Fax +917926857321				
Website: www.nprocure.com				

(Dr. Shaliesh Arlekar)
I/c. Medical Superintendent,
Government Hospital, Daman

**ADMINSTRATION OF DAMAN AND DIU,
O/O. THE MEDICAL SUPRINTENDENT,
GOVERNMENT HOSPITAL, DAMAN.**

TERMS & CONDITIONS FOR THE

**“Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for
Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for
Government Hospital, Daman”**

E-Tender Notice No.GHD/E-TENDER/2016-2017/495

Dated. 20/07/2016

1.	The rate(s) quoted should be strictly for free delivery at F.O.R. Government Hospital, Daman and will be valid and operative for supply orders issued on or before 31-03-2017 and rate should not be more than MRP. Valid for 1 year from date of Issue.	
2.	All Taxes/ Duties/ Royalties charges payable on sales/transport etc. within and / or outside the UT / State shall be payable by the supplier.	
3.	No extra charge for packing, forwarding and insurance etc. will be paid on the rates quoted.	
4.	The rates should be quoted only for the items specified in the list of requirement and should be for the items of given special mark / manufacture.	
5.	Rates quoted for items other than required specification/mark/manufacture may not be considered. However, indigenous manufactures may quote their own makes provided the specifications/mark/manufacture.	
6.	Where specification / mark/ manufacture are not specifying by this office, the rates should be quoted only for the first class and standard quality only.	
7.	The tenderer should specify the name of the manufacturer for the item quoted by him along with catalogue of the item.	
8.	The decision of the E-Tender Inviting Officer for acceptance/rejection of “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” supplied including the decision for equivalent specifications, standard and quality etc.shall be final.	
9.	The tenderer should send in advance or enclose along with technical bid an amount of Rs.1,35,000/- as Earnest Money Deposit (FDR) in form of Nationalized of any Bank payable at Daman in favour of the officer inviting tender i.e. Medical Superintendent, Government Hospital, Daman. The EMD submitted other than form mentioned above will not be accepted. Tender received without Earnest Money Deposit will be summarily rejected.	
10.	(a)	The successful tenderer will have to pay within 10 days from the date of demand, an amount equal to 10% of the total value of articles, that may be order, as the amount of Security Deposit.
	(b)	Non-receipt of Security Deposit within stipulated time limit will result in automatic cancellation of the order for supply without any intimation.
	(c)	However in case if any articles are received for which the Security Deposit may not have been deposited, the full Security Deposit as may be due from the supplier will be recovered from the bills for such articles.
11.	The amount of Earnest Money paid by the successful Tenderer(s) will be adjusted against the amount of Security Deposit to be paid by the successful tenderer(s) as per condition No. 10 above.	
12.	All bills for amount above Rs.5000/- should be pre-receipted on a Revenue Stamp of proper value. Bills for amount exceeding Rs.5000/- not pre-receipted on Revenue Stamp of proper value will not be accepted for payment.	
13.	Each bill in which Sales Tax/Vat is charged must contain the following certificate on the body of the bill. "CERTIFIED that the goods on which Sales Tax/VAT has been charged have not been exempted under the VAT/ Central Sale Tax Act or the Rules made there under and the amount charged on account of VAT/Sales Tax on these goods is not more than what is payable under the provisions of relevant Act or Rules made there under".	
14	The right to accept or reject without assigning any reasons or all tenders in part or whole is reserved with the E-Tender Inviting Officer and his decision(s) on all matters relating to acceptance or rejection of the tenders as a whole or in part will be final and binding to all	

15	If the tenderer whose tender is accepted, fails to execute the supply order within stipulated time the Earnest Money Deposit of such tenders will stand forfeited to the Government.	
16	In case, the supplier does not execute the supply order placed with him, the EMD of the supplier will be forfeited to the Government and the contract for the supply shall terminated with no further liabilities on either party to the contract.	
17	No separate agreement will be required to be signed by the successful tender(s) for the purpose of this contract for supply. Rates tendered/offered in response to the concerned Tender Notice shall be considered as acceptance of all above terms and conditions for supply for all legal purpose.	
18	The E-Tender should be neatly typed only on letterhead carry the name of supplier and the signature of the tenderer. No overwriting correction or erasures will be considered.	
19	The rates quoted should be inclusive of all taxes, duties, surcharges, cess, freight, loading, unloading, insurance, road permits, packing, (VAT and other taxes if applicable) as applicable. No extra taxes will be given.	
20	All bills should be in TRIPLICATE and should invariably mention the number and date of supply order.	
	The Tenders and financial bid should be submitted online on www.nprocure.com in two bid system.	
21.	Orders once placed should be delivered within the given time period and item should be door delivered.	
22.	No extra charge for packing, forwarding and insurance etc. will be paid on the rates quoted.	
	The rates should be quoted only for the items specifies in the list of requirement and should be for the items of given special mark/manufacture only.	
23	Rates quoted for items other than the required specification / mark / manufacture will not be considered.	
24	The amount of Earnest Money paid by the tenderer(s) whose tenders are not accepted will be refunded to them.	
25	Only on satisfactory completion of the supply order for and on payment of all bills of the supplier, as to be admitted for payment, the amount of Security Deposit/Earnest Money will be refunded after expiry of guarantee/warranty period, if any, or any such date/period as may be mutually agreed upon	
26	The E-tender inviting officer will consider extension of time for remitting the Security Deposit as demanded. However, in case of denial to consider such extension the supplier is bound to abide by the limit given and liable to make good any loss to the Government on account of his failure to abide by the time limit.	
27	(a)	Railway Receipt or other transport document should be drawn in favor of the Officer Inviting Tenders.
	(b)	Railway Receipt or other transport document should not be send by V.P.P. or through any Bank as this being a Government Office it is not possible to clear cash demands of Post Office/Bank for delivery of R.R. or other transport documents unless we have agreed to it as special arrangements.
	(c)	Railway Receipt or transport receipt should be sent to this Office by Registered Post immediately on dispatch of goods from dispatching end.
28	The supplies of “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” of inferior quality standard or of different specifications, doses/content, brand, manufacturer etc. other than that ordered specified and/or incomplete or broken articles will not be accepted. The supplier has to replace the same at his own cost and risk. Intimation of non-acceptance of any “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” will be sent to the supplier within 10 days from the date of receipt of the stores and the same will be returned to the supplier at his own cost and risk, if he so desires and intimates accordingly within 15 days from the date of dispatch of intimation of the non-acceptance. However, if no communication is received within 15 days from the date of communication the tender Inviting Officer will not be responsible for any damage, loss etc. of such rejected articles.	
29.	In case of failure to replace the accepted and rejected articles from supplies made, as mentioned in the conditions, the loss undergone by the Government	

	will be recovered from the suppliers Security Deposit/Earnest Money or payment due of any bill(s) to the extent required.
30	In case of failure to Purchase of “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” ordered for, as per conditions and within the stipulated time, the same will be obtained from the tenderer who offered next higher rates or from any other sources, as may be decided by the tender inviting Officer and the loss to the Government on account of such purchases(s) shall be recovered from the former suppliers Security Deposit/Earnest Money or bills payable. The suppliers shall have no any right to dispute with such procedure.
31.	Extension of time limit for supplies shall be consider by the Tender Inviting Officer. The extension so granted may be with levy of compensation for delay in execution of supply order up to 5% of the cost of supplies ordered for at the discretion of the authority competent to grant extension of time limit provided such request is made well in time, depending upon the circumstances and such decision in the matter will be final.
	Demurrage charges paid by the E-Tender Inviting Officer on account of delayed receipt of dispatch documents intimation will be recovered from the bills payable to the supplier.
32	If at any time after the order for “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” the E-Tender Inviting Officer shall for any reason whatsoever not require the whole or part of the quantity thereof as specified in the order the tender Inviting Officer shall give notice in writing of the fact to the supplier(s) who shall have no claim to any payment of compensation what so ever on account of any profit or advantage which the supplier(s) might have derived from the supply of articles in full, but which did not derive in consequence of the full quantity of articles not having been purchased, nor shall have any claim for compensation by reasons of any alterations having been made in the original instructions which shall invoice any curtailment of the supply originally contemplated.
33	Rates should be quoted as per the forms prescribed by the department and as per the requirement asked for.
34	The Earnest Money(s)/Security Deposit(s) paid by the tender(s) earlier against any tender(s) or supply order(s) is/are not adjustable with Earnest Money or Security Deposit required by these conditions.
35	The Tenders/offers received which do not confirm with the terms and conditions of this office will be summarily rejected. If any firm desires to consider exemption from payment of Earnest Money Deposit, certified copies of its Registration with D.G.S. & D. should be attached to their tenders.
36	Supplier may ensure the goods at his own cost to safeguard the delivery of such goods dispatched by him to the consignee; the department will not be responsible for the damage or pilferage of goods during transit.
37	The tenderer should attached Scanned copies of Certificate from the respective manufacture company indicating your firm as authorized dealer/supplier for their product, certificate of experience in the field of “Electrical Operation Table” valid license, proof of fulfilling the norms of ISI/ISO/WHO/GMP, CE Certified specification if any, copy of dealership letter, licence for import, PAN No, Sales Tax No., The supplier must have at least 03 years experience of supplying such good and shall average annual turnover of Rs.13,50,000/- Suring last three years ending by 31.12.2016 A letter indicating last three financial year Turnover figures and profit & Loss Accounts should accompany with Technical Bid for Verification. Please note that the balance sheets should carry average annual financial turnover during last 3 years up to 31/12/2016 and should be at least 30% of the estimated cost.
38	Rates quoted are for “Purchase and Installation of 03 Nos. of Electrical Operation Table with Central Pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman” to be supplied at Government Hospital, Daman.
39	Tenderer should enclose along with tender an amount of Rs.1,35,000/- as Earnest Money Deposit in form of Fixed Deposit Receipt of any Nationalized Bank payable at Daman in favour of Medical Superintendent, Government

	Hospital, Daman. The EMD Submitted other than Form mentioned above will not be accepted. Tender received without EMD will be summarily rejected.
40	The tender fee must be enclosed in demand draft in favour of undersigned with the tender documents.
41	AMC: The rates of Annual Maintenance Contract (AMC) should be for the period of three years from the expiry of Guarantee / Warranty period of one year and should written separately in the Financial bid year-wise. The decision to accept the tender with or without AMC is reserve with the Purchase Committee.
42	The tenderers should give the guaranty / warranty for the period of not less than one year from the date of installation against any manufacturing defect.
43	The tenderer may be called for a sample / demonstration of the items quoted for which they will be informed one week in advance for arranging the necessary sample / demonstration in the hospital on a suitable date and time failing which the tender will be rejected.
44	The tenderers shall be bound to give assurance for undertaking the Annual Maintenance Contract after expiry of guarantee / warranty period.
45	The successful tenderers shall be bound to provide training if any required without any extra charges during commissioning.
46	The successful tenderers should install and commissioning the equipment at the site suggested by the office.
47	The tendered quantity is tentative and the actual purchase can be 15% less or more than the quantity put to tender for all items and the tenderer is bond to supply such requirement without any demur.
48	The tendering firm must be registered with the Sales Tax/ VAT department and a copy of their registration under the Sales Tax/ VAT bearing the TIN Number be provided.
	The Tender should give the guaranty/warranty for the period of not less than one year from the date of installation against any manufacturing defect.
49	The terms and conditions dully accepted and signed by the tenderer should attached with the Technical Bid.
50	The Financial Bid has to fill in prescribed format.
51	Tender should furnished a non black listed certificate that the firm has not been black listed in the past by any Government/private institution that there is no vigilance / CBI case against the firms supplier and the firm has not been blacklisted in the past by any Government or private organization, the tenderers should be attached an undertaking. (Annexure-I)
52.	The bidder should not have quoted price lower than quoted in this tender to any Govt./Semi Govt./public sector undertaking, bidder should enclosed affidavit to that effect on a non judicial stamp paper of Rs.20/-

Signature & Designation of
Tender Inviting Officer

(Dr. Shailesh Arlekar)
I/c. Medical Superintendent,
Daman.

The above terms and conditions are accepted and are binding to me / us.

Place :

Signature of Tenderers

Dated :

Name of Tenderers with seal
of the firm

NOTE: Please return one copy of these terms and conditions dully sign with seal of firm along with the tender.

ANNEXURE – I

(to be submitted on the letter head of the company/firm)

I hereby certify that the above firm has not been ever blacklisted by any Central/State Government/Public Undertaking/Institution on any account.

I also certify that the above information is true and correct in any every respect and in any case at a later date it is found that any details provided above are incorrect, any contract given to the above firm may be summarily terminated and the firm blacklisted.

I also certify that firm will supply the item as per the specifications given by institution and also abide all the terms and conditions stipulated in tender.

Date:

Place:

Name:

Business Address:

Signature Bidder:

Seal of Bidder:

**ADMINISTRATION OF DAMAN AND DIU,
O/O. THE MEDICAL SUPERINTENDENT,
GOVERNMENT HOSPITAL, DAMAN.**

E-Tender Notice No.GHD/E-TENDER/2016-2017/495

Dated. 20/07/2016

ELIGIBILITY PARAMETERS.

1.	Name of the Organization	
2.	Postal Address	
3.	Telephone / Fax / Mobile No. of the Organization / Firm.	
4.	Status of the Organization/Firm (whether Private or Public Sector undertaking or Sole Proprietor or Partnership or co-operative society etc.) The tender should attach a resolution passed by the Executive Body authorizing the specific officer / partner for signing the documents.	
5.	Certificate from the respective manufacture company indicating your firm as authorized dealer/supplier for their product	
6.	Certificate of experience in the field of “Purchase and Installation of 03 Nos. of Electrical Operation Table with central pillar for Gaynec Surgery, General Surgery and Neuro and Orthopedic Surgery with accessories for Government Hospital, Daman”	
7.	Valid licence	
8.	Proof of fulfilling the norms of ISI/ISO/WHO/GMP	
9.	US FDA, CE and ISO Certified specified specification, if any	
10.	Copy of dealership letter	
11.	Licence for import	
12.	PAN CARD No.	
13.	Sales Tax / VAT No.	
14.	Service Tax if any	
15.	Copy of Turnover for Rs. 13,50,000/-of the estimated cost during last 3 years up to 31/12/2016	

Place :

Date :

Signature of Tenderer
Name of Tender with
Seal of the firm

**ADMINISTRATION OF DAMAN AND DIU
O/O. THE MEDICAL SUPRINTENDENT
GOVRNMENT HOSPITAL, DAMAN**

Tender From (Technical Bid)
No.GHD/E-TENDER/2016-2017/495 Dated: 20/07/2016.

1	<u>SPECIFICATIONS FOR ELECTRICAL TABLE WITH CENTRAL PILLAR FOR GAYNEC SURGERY</u>	Company/ Model No.	Remark
1	Actuator operator, Eight Function remote control table.		
	• Up & Down (Min-28", Max-40").		
	• Trendelenburg & Reverse Trendelenburg (30° either side).		
	• Right Lateral & Left Lateral (21° either side).		
	• Back rest- Motorized (+60° to -40°).		
2	Power supply should be is 220 v, 5 amps and three pin Domestic Plug.		
3	It should be mounted on four imported polyurethane castors for longitudinal & lateral movements.		
4	Braking should be very effective and has a dead lock when braked; it is on the head end side of the table, to facilitate the anesthetist,		
5	Leg bed has two pieces right and left and should be detachable and abductable and also can be moved up & down.		
6	Table top has five sections; it should be breakable into inverted v-shape from the head side.		
7	Trendelenberg and lateral mechanism should be concealed with bellows; column should be Straight without any horizontal projection on right or left.		
8	The up & down movements should jerk free & mounted with high precision glides to totally avoid wig-wag movement in the column.. The column should be made out of steel and not aluminium. Because steel is more stronger compared to aluminium		
9	Table column size should very sleek for 95° 'C'-arm access.		
10	Polyurethane detachable moulded cushion top without joints and stitches should be provided on the table top.		
11	Base, column casing, table top frame, traction bars and all accessories should be made of non-corrosive steel.		
12	The table top should be made of 50x50 mm square pipe , 10 swg thickness		
13	The table top should be reversible, the headrest attachments can be fixed to the tail side and the tail side attachment can be fixed to the head side.		
14	Head rest should be detachable.		
15	The backrest should be operated electrically through remote control.		
16	Patient sitting position, with Trendelenburg maximum should be possible.		
17	All the actuators used in the table should be of 10000 N, and it should be used as a whole actuator and not assembled here with gear motor and self designed screw rod		
18	The bottom of the table should be sealed with rust proof material to prevent water and dust from entering inside		
19	There should be a provision to incorporate "0" function if required		
20	Should be C.E, TUV-ISO-9001 AND ITC-EN ISO-13485/2003 Approved.		
21	The table should have facility of Manual override system if there is any fault or breakdown in electro hydraulic system.		
22	All accessories clamps should be die casted S.S 304 grade		
23	There should a facility for Zero positioning from any position at a single push of button.		
24	Die presses D.S 304 grade cover on base without any edges and welding.		

25	Head & leg sections should be interchangeable		
26	It should have Electric floor locking mechanism		
27	Very heavy duty base for total stability		
28	Positions by remote:-		
	· Up – Down		
	· Trendelenburg & Rev. Trendelenburg		
	· Lateral support Tilt		
	· Flex – Reflex		
	· Electrical Floor Locking		
29	Following Standard Accessories should be supplied with the table:		
	· Anesthetic Screen		
	· Shoulder Support with pads		
	· Arm Boards with pads		
	· Lateral support with pad		
	· Knee Crutches Gospel type		
	· Water Proof Rubber mattress		
	· Accessories Stand made of Stainless steel 304		
	The accessories clamp should be made of 304 G S.S die casted with any welding. All the bright part should be made of 304 G St. Steel. It must have CE and ISO Certification.		
2	<u>SPECIFICATIONS FOR ELECTRICAL TABLE WITH CENTRAL PILLAR FOR GENERAL AND LAPROSCOPY SURGERY</u>		
1	Actuator operator, Eight Function remote control table.		
	· Up & Down (Min-28”, Max-40”).		
	· Trendelenburg & Reverse Trendelenburg (30° either side).		
	· Right Lateral & Left Lateral (21° either side).		
	· Back rest- Motorized (+60° to –40°).		
2	Power supply should be is 220 v, 5 amps and three pin Domestic Plug.		
3	It should be mounted on four imported polyurethane castors for longitudinal & lateral movements.		
4	Braking should be very effective and has a dead lock when braked; it is on the head end side of the table, to facilitate the anesthetist,		
5	Leg bed has two pieces right and left and should be detachable and abductable and also can be moved up & down.		
6	Table top has five sections; it should be breakable into inverted v-shape from the head side.		
7	Trendelenberg and lateral mechanism should be concealed with bellows; column should be straight without any horizontal projection on right or left.		
8	Entirely off centered table top should provides unrestricted ‘C’ arm imaging for entire body.		
9	The up & down movements should jerk free & mounted with high precision glides to totally avoid wig-wag movement in the column.. The column should be made out of steel and not aluminium. Because steel is more stronger compared to aluminium		
10	Table column size should very sleek for 95° ‘C’-arm access.		
11	Table should have the provision to view spine AP and lateral with ‘C’-arm.		
12	Polyurethane detachable moulded cushion top without joints and stitches should be provided on the table top.		
13	Base, column casing, table top frame, traction bars and all accessories should be made of non-corrosive steel.		
14	The table top should be made of 50x50 mm square pipe , 10 swg thickness		

	The table top should be reversible, the headrest attachments can be fixed to the		
15	tail side and the tail side attachment can be fixed to the head side.		
16	Head rest should be detachable.		
17	The backrest should be operated electrically through remote control.		
18	The entire vertebral column should be viewed without any hindrance.		
19	Patient sitting position, with Trendelenburg maximum should be possible.		
20	It should adapt Mayfield sugitha and Lyela retractor.		
21	‘C’arm compatible pelvic surgeries should be performed.		
22	PCNL surgery should be possible without any extra floor support		
	All the actuators used in the table should be of 10000 N, and it should be used		
23	as a whole actuator and not assembled here with gear motor and self designed screw rod		
	The bottom of the table should be sealed with rust proof material to prevent		
24	water and dust from entering inside		
25	There should be a provision to incorporate “0” function if required		
26	Should be C.E, TUV-ISO-9001 AND ITC-EN ISO-13485/2003 Approved.		
	The table should have facility of Manual override system if there is any fault or		
27	breakdown in electro hydraulic system.		
28	All accessories clamps should be die casted S.S 304 grade		
	There should a facility for Zero positioning from any position at a single push of		
29	button.		
30	Die presses D.S 304 grade cover on base without any edges and welding.		
31	Head & leg sections should be interchangeable		
32	It should have Electric floor locking mechanism		
33	Very heavy duty base for total stability		
34	Positions by remote:-		
	· Up – Down		
	· Trendelenburg & Rev. Trendelenburg		
	· Lateral support Tilt		
	· Flex – Reflex		
36	· Electrical Floor Locking		
	Following Standard Accessories should be supplied with the table:		
	· Anesthetic Screen		
	· Shoulder Support with pads		
	· Arm Boards with pads		
	· Lateral support with pad		
	· Knee Crutches Gospel type		
	· Water Proof Rubber mattress		
	· Accessories Stand made of Stainless steel 304		
	The accessories clamp should be made of 304 G S.S die casted with any welding.		
	All the bright part should be made of 304 G St. Steel. It must have CE and ISO Certification.		
	<u>Standard Accessories for General Surgery</u>		
S.NO	ITEM	QTY	
1	Arm rest	2 Nos	
2	Simple side end block	3 Nos	
3	Extension bar	2 Nos	
4	Lithotomy	2 Nos	
5	Side Support with clamp	1 No	
6	Anesthesia screen rod	1 No	
7	Radial setting clamp	2 Nos	
8	Shoulder support with clamp	2 Nos	
	<u>Optional Accessories for General Surgery:</u>		
S.NO	ITEM	QTY	

1	Short Uro/gyn bed	1 Nos	
2	Bolsters	2 Nos	
3	Incorporated kidney bridge	1 Nos	
4	Battery Back up	1 Nos	
5	Zero function	1 Nos	
3	<u>Specification for Electrical Table with Eccentric Pillar for Neuro and Orthopedic</u>		
1	Actuator operator, Eight Function remote control table.		
	• Up & Down (Min-28”, Max-40”).		
	• Trendelenburg & Reverse Trendelenburg (30° either side).		
	• Right Lateral & Left Lateral (21° either side).		
	• Back rest- Motorized (+60° to –40°).		
2	Power supply should be is 220 v, 5 amps and three pin Domestic Plug.		
3	It should be mounted on four imported polyurethane castors for longitudinal & lateral movements.		
4	Braking should be very effective and has a dead lock when braked; it is on the head end side of the table, to facilitate the anesthetist,		
5	Leg bed has two pieces right and left and should be detachable and abductable and also can be moved up & down.		
6	Table top has five sections; it should be breakable into inverted v-shape from the head side.		
7	Trendelenberg and lateral mechanism should be concealed with bellows; column should be Straight without any horizontal projection on right or left.		
8	Entirely off centered table top should provides unrestricted ‘C’ arm imaging for entire body.		
9	The up & down movements should jerk free & mounted with high precision glides to totally avoid wig-wag movement in the column.. The column should be made out of steel and not aluminium. Because steel is more stronger compared to aluminium		
10	Table column size should very sleek for 95° ‘C’-arm access.		
11	Table should have the provision to view spine AP and lateral with ‘C’-arm.		
12	Polyurethane detachable moulded cushion top without joints and stitches should be provided on the table top.		
13	Base, column casing, table top frame, traction bars and all accessories should be made of non-corrosive steel.		
14	The table top should be made of 50x50 mm square pipe , 10 swg thickness		
15	The table top should be reversible, the headrest attachments can be fixed to the tail side and the tail side attachment can be fixed to the head side.		
16	Head rest should be detachable.		
17	The backrest should be operated electrically through remote control.		
18	The entire vertebral column should be viewed without any hindrance.		
19	Patient sitting position, with Trendelenburg maximum should be possible.		
20	It should adapt Mayfield sugitha and Lyela retractor.		
21	‘C’arm compatible pelvic surgeries should be performed.		
22	PCNL surgery should be possible without any extra floor support		
23	All the actuators used in the table should be of 10000 N, and it should be used as a whole actuator and not assembled here with gear motor and self designed screw rod		
24	The bottom of the table should be sealed with rust proof material to prevent water and dust from entering inside		
25	There should be a provision to incorporate “0” function if required		
26	Should be C.E, TUV-ISO-9001 AND ITC-EN ISO-13485/2003 Approved.		

27	The table should have facility of Manual override system if there is any fault or breakdown in electro hydraulic system.		
28	All accessories clamps should be die casted S.S 304 grade		
29	There should a facility for Zero positioning from any position at a single push of button.		
30	Die presses D.S 304 grade cover on base without any edges and welding.		
31	Head & leg sections should be interchangeable		
32	It should have Electric floor locking mechanism		
33	Very heavy duty base for total stability		
34	Positions by remote:-		
	· Up – Down		
	· Trendelenburg & Rev. Trendelenburg		
	· Lateral support Tilt		
	· Flex – Reflex		
35	· Electrical Floor Locking		
	Following Standard Accessories should be supplied with the table:		
	· Anesthetic Screen		
	· Shoulder Support with pads		
	· Arm Boards with pads		
	· Lateral support with pad		
	· Knee Crutches Gospel type		
	· Water Proof Rubber mattress		
	· Accessories Stand made of Stainless steel 304		
	The accessories clamp should be made of 304 G S.S die casted with any welding. All the bright part should be made of 304 G St. Steel. It must have CE and ISO Certification.		
	<u>Standard Accessories for Neuro Table</u>		
S.NO	ITEM	QTY	
1	Arm rest	2 Nos	
2	Side Support with clamp	1 Nos	
3	Anesthesia screen rod	1 Nos	
4	Simple side end block	3 Nos	
5	Adaptor for shoulder arthroscopy and cervical attachment	1 Nos	
6	Horse shoe shaped cervical attachment	1 Nos	
7	Spine extension plate	1 Nos	
8	Neuro sitting position	1 Nos	
9	Bolsters	2 Nos	
	Standard Accessories for Orthopedic Table:		
S.NO	ITEM	QTY	
1	Arm rest	2 Nos	
2	Simple side end block	3 Nos	
3	Extension bar	2 Nos	
4	Lithotomy	2 Nos	
5	Side Support with clamp	1 Nos	
6	Anesthesia screen rod	1 Nos	
7	Radial setting clamp	2 Nos	
8	Shoulder support with clamp	2 Nos	
9	Orthopaedic traction attachment (for DHS / femur) (single joint)	1 set	
10	DHS/femur counter traction post	1 Nos	
11	Telescopic traction right	1 Nos	
12	Telescopic traction left	1 Nos	
13	DHS Extension	1 Nos	
14	Left traction boot	1 Nos	
15	Right traction boot	1 Nos	
16	Screw tension device	2 Nos	
17	Hand surgery bed	1 Nos	
18	Humorous bed	1 Nos	

OPTIONAL Accessories for Neuro Table

S.NO	ITEM	QTY	
1	Four post spinal frame	1 Nos	
2	Bow shaped spinal frame	1 Nos	
3	Radiolucent bow shaped spinal frame	1 Nos	
4	Three pin Fixator	1 Nos	
5	Four pin Fixator	1 Nos	
6	Leyla retractor	1 Nos	
7	Battery Back up	1 Nos	
8	Zero function	1 Nos	

Optional accessories for Orthopedic Table:

S.NO	ITEM	QTY	
1	Tibial attachment with thigh holders and telescopic extension for traction	1 set	
2	Steinman pin holder	1 Nos	
3	Floor support	2 Nos	
4	Adaptor for shoulder arthroscopy and cervical attachment	1 Nos	
5	Horse shoe shaped cervical attachment	1 Nos	
6	THR/TKR attachment	1 Nos	
7	Spine extension plate	1 Nos	
8	Bolsters	2 Nos	
9	Short Uro/gyn bed	1 Nos	
10	Incorporated kidney bridge	1 Nos	
11	Accessories stand	1 Nos	
12	Battery Back up	1 Nos	
13	Zero function	1 Nos	

ADMINISTRATION OF DAMAN AND DIU
O/O. THE MEDICAL SUPRINTENDENT
GOVERNMENT HOSPITAL, DAMAN
Tender From (Financial Bid)

No.GHD/E-TENDER/2016-2017/495

Dated: 20/07/2016.

SR NO.	NAME OF ITEMS	Unit	Rate in figure	Rate in words
(A)	Rate as per Technical Specification Electrical Table with Central Pillar for Gynec Surgery with accessorias			
1	Anesthetic Screen	01 Nos.		
2	Shoulder Support with pads	01 Nos.		
3	Arm Board with pads	01 Nos.		
4	Lateral Support with pad	01 Nos.		
5	Knee crutes Gospel type	01 Nos.		
6	Water Proof Rubber mattress	01 Nos.		
7	Accessories stand mad of stainless steels 304	01 Nos.		
(A) (i)	Electrical Table with Central Pillar for Gynec Surgery with accessorias			
i	First Year	1		
ii	Second Year	1		
iii	Third Year	1		
iv	Forth Year	1		
v	Fifth Year	1		
B	Rate as per Technical Specification Electrical Table with Central Pillar for General and Laproscopy Surgery with accessorias			
1	Arm rest	02 Nos.		
2	Simple side end block	03 Nos.		
3	Extraction bar	02 Nos.		
4	Lithotomy	02 Nos.		
5	Side Support with clamp	01 Nos.		
6	Anesthesia screen rod	01 Nos.		
7	Radial setting clamp	02 Nos.		
8	Shoulder support with clamp	02 Nos.		
9	Short Uro/gyn bed	01 Nos.		
10	Bolsters	02 Nos.		
11		01 Nos.		

i	First Year	1		
ii	Second Year	1		
iii	Third Year	1		
iv	Forth Year	1		
v	Fifth Year	1		
(C)	Rate as per Technical Specification Electrical Table with Eccentric Pillar for Neuro and Orthopedic Surgery with accessorieas			
1	Arm rest	02 Nos.		
2	Side Support with clamp	01 Nos.		
3	Anesthesia screen rod	01 Nos.		
4	Simple side end block	03 Nos.		
5	Adaptor for shoulder arthroscopy and cervical attachment	01 Nos.		
6	Horse shoe shaped cervical attachment	01 Nos.		
7	Spine extension plate	01 Nos.		
8	Neuro sitting position	01 Nos.		
9	Bolsters	02 Nos.		
10	Four post spinal frame	01 Nos.		
11	Bow shapped spinal frame	01 Nos.		
12	Radiolucent bow shaped spinal frame	01 Nos.		
13	Three pin Fixator	01 Nos.		
14	Four pin Fixetor	01 Nos.		
15	Leyla retractor	01 Nos.		
16	Battery Back up	01 Nos.		
17	Zero function	01 Nos.		
18	Arm rest	02 Nos.		
19	Simple side end block	03 Nos.		
20	Extension bar	02 Nos.		
21	Lithotomy	02 Nos.		
22	Side support with clamp	01 Nos.		
23	Anesthesia screen roe	01 Nos.		
24		02 Nos.		

28	Telescopic traction right	01 Nos.		
29	Telescopic traction left	01 Nos.		
30	DHS Extension	01 Nos.		
31	Left traction boot	01 Nos.		
32	Right traction boot	01 Nos.		
33	Screw tension device	02 Nos.		
34	Hand Surgery bed	01 Nos.		
35	Humerous bed	01 Nos.		
36	Tibial attachment with thigh holders and telescopic extension for traction	01 Set		
37	Steinman pin holder	01 Nos.		
38	Floor support	01 Nos.		
39	Adaptor for shoulder arthroscopy and cervical attachment	01 Nos.		
40	Horse shoe shaped cervical attachment	01 Nos.		
41	THR/ TKR attachment	01 Nos.		
42	Spine extension plate	01 Nos.		
43	Bolster	02 Nos.		
44	Short Uro/ gyn bed	01 Nos.		
45	Incorporated kidney bridge	01 Nos.		
46	Accessories stand	01 Nos.		
47	Battery Back up	01 Nos.		
48	Zero function	01 Nos.		
C (i)	Electrical Table with Eccentric Pillar for Neuro and Orthopedic Surgery with accessories			
i	First Year	1		
ii	Second Year	1		
iii	Third Year	1		
iv	Forth Year	1		
v	Fifth Year	1		